
Vitronectin (VTN) Assay Kit for Rat

PRODUCT OVERVIEW

Product Name	Vitronectin (VTN), Assay Kit
Full Product Name	Vitronectin (VTN) Magnetic Fluorescence Assay Kit
Product Gene Name	VTN assay kit
Product Synonym Names	V75; VN; VNT; Serum Spreading Factor; Somatomedin B; Complement S-Protein; Vitronectin
Species Reactivity	Rat
Assay Type	Double-antibody Sandwich
Samples	Serum, plasma, tissue homogenates, cell lysates, cell culture supernates and other biological fluids
Detection Range	0.1-100ng/mL
Sensitivity	The minimum detectable dose of this kit is typically less than 0.033 ng/mL
Specificity	This assay has high sensitivity and excellent specificity for detection of Vitronectin (VTN), etc. No significant cross-reactivity or interference between Vitronectin (VTN), etc. and analogues was observed.
Preparation and Storage	The stability of kit is determined by the loss rate of activity. The loss rate of this kit is less than 5% within the expiration date under appropriate storage condition. To minimize extra influence on the performance, operation procedures and lab conditions, especially room temperature, air humidity, incubator temperature should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same operator from the beginning to the end.
Supply Chain Verification	Manufactured in a lab with traceable raw materials. Bulk orders can typically be prepared to the customer's specifications, please inquire.
Precision	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level Vitronectin (VTN), etc. were tested 20 times on one plate, respectively. Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level Vitronectin (VTN), etc. were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD/meanX100 Intra-Assay: CV<10% Inter-Assay: CV<12%
Research Area	Signal transduction, Metabolic pathway, Tumor immunity, Hepatology
Assay Procedure Summary	1. Preparation of standards, reagents and samples before the experiment 2. Add 100uL standard or sample to each well, add 10uL magnetic beads, and incubate 90min at 37 degree C on shaker 3. Remove liquid on magnetic frame, add 100uL prepared Detection Reagent A. Incubate 60min at 37 degree C on shaker 4. Wash plate on magnetic frame for three times 5. Add 100uL prepared Detection Reagent B, and incubate 30 min at 37 degree C on shaker 6. Wash plate on magnetic frame for three times 7. Add 100uL sheath solution, swirl for 2 minutes, read on the machine.
Test Principle	Analyte-specific antibodies are pre-coated onto color-coded microparticles. Microparticles, standards, and samples are pipetted into wells and the immobilized antibodies bind the analytes of interest. After washing away any unbound substances, a biotinylated antibody cocktail specific to the analytes of interest is added to each well. Following a wash to remove any unbound biotinylated antibody, Streptavidin-Phycoerythrin conjugate (Streptavidin-PE), which binds to the biotinylated detection antibodies, is added to each well. A final wash removes unbound Streptavidin-PE and the microparticles are resuspended in buffer and read using fluorescent spectrum analyzer.The MFI developed isproportional to the concentration of analytes of interest in the sample.
ISO Certification	Manufactured in an ISO 9001:2015 and ISO 13485:2016 Certified Laboratory.

Application Notes for VTN assay kit	Magnetic Fluorescence Assay for Antigen Detection.
Related Product Information for VTN assay kit	MBS2156842 is an assay kit targeting VTN. Also known as Vitronectin (VTN) Magnetic Fluorescence Assay Kit. Reactive in Rat.
Other Notes	Small volumes of VTN assay kit vial(s) may occasionally become entrapped in the seal of the product vial during shipment and storage. If necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the container`s cap. Certain products may require to ship with dry ice and additional dry ice fee may apply.



Typical Testing Data/Standard Curve (for reference only) of VTN assay kit

ADDITIONAL DETAILS

NCBI/Uniprot data below describe general gene information for VTN. It may not necessarily be applicable to this product.

Protein Family	Vitronectin
Precautions	All of MyBioSource's Products are for scientific laboratory research purposes and are not for diagnostic, therapeutics, prophylactic or in vivo use. Through your purchase, you expressly represent and warrant to MyBioSource that you will properly test and use any Products purchased from MyBioSource in accordance with industry standards. MyBioSource and its authorized distributors reserve the right to refuse to process any order where we reasonably believe that the intended use will fall outside of our acceptable guidelines.
Disclaimer	While every efforts were made to ensure the accuracy of the information provided in this datasheet, MyBioSource will not be liable for any omissions or errors contained herein. MyBioSource reserves the right to make changes to this datasheet at any time without prior notice. It is the responsibility of the customer to report product performance issues to MyBioSource within 30 days of receipt of the product. Please visit our Terms & Conditions page for more information.
Product Note	Select online data sheet information is drawn from bioinformatics databases, occasionally resulting in ambiguous or non-relevant product information. It is the responsibility of the customer to review, verify, and evaluate the information to make sure it matches their requirements before purchasing the kit. Our ELISA Kit assays are dynamic research tools and sometimes they may be updated and improved. If the format of this assay is important to you then please request the current manual or contact our technical support team with a presales inquiry before placing an order. We will confirm the current details of the assay. We cannot guarantee the sample manual posted online is the most current manual, it is intended to serve as an example only. Please refer to the instructions For Use provided with the assay kit for precise details.

For Research Use Only. Not for use in diagnostic or therapeutic procedures.

MyBioSource | Tel: +1.858.633.0165 | Fax: +1.858.633.0166 | mybiosource.com/contact-us
Online datasheet: https://www.mybiosource.com/rat-assay-kits/vitronectin-vtn/2156842