Copper Transporter 1 (COPT1) ELISA Kit for Human

PRODUCT OVERVIEW

Product Name	Copper Transporter 1 (COPT1), ELISA Kit
Full Product Name	Human Copper Transporter 1 ELISA Kit
Product Gene Name	COPT1 elisa kit
Species Reactivity	Human
Assay Type	Sandwich
Samples	Serum, Plasma, Tissue Homogenate, Feces and Urine
Detection Range	3.12ng/ml-100ng/ml.
Sensitivity	1.0ng/ml.
Specificity	No significant cross-reactivity or interference between Human COPT1 and analogues was observed.
Preparation and Storage	Store all reagents at 2-8 degree C
Inter-assay Precision	Inter-assay CV (%) is less than 15%. [CV(%) = SD/mean ×100]
Intra-assay Precision	Intra-assay CV (%) is less than 15%.
Intended Uses	This Quantitative Sandwich ELISA kit is only for in vitro research use only, not for drug, household, therapeutic applications! It is intended to be determinated COPT1 concentrations in Human serum, plasma and other body fluids. Using Purified Human COPT1 antibody to coat Microelisa Stripplate wells to make solid-phase antibody, then add COPT1 and COPT1 antibody which has been labeled with HRP to wells, then the reactants become antibody-antigen-antibody-enzyme complex, after washing completely, add TMB substrate solution, TMB substrate becomes blue color under HRP enzyme-catalyzed, reaction is terminated by the addition of a sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450 nm. The concentration of COPT1 in the samples is then determined by comparing the O.D. of the samples to the standard curve.
Related Product Information for COPT1 elisa kit	**Background:** This Quantitative Sandwich ELISA kit is only for in vitro research use only! It is intended to be determinated COPT1 concentrations in Human serum, plasma and other body fluids. Using Purified Human COPT1 antibody to coat Microelisa Stripplate wells to make solid-phase antibody, then add COPT1 and COPT1 antibody which has been labeled with HRP to wells, then the reactants become antibody-antigen-antibody-enzyme complex, after washing completely, add TMB substrate solution, TMB substrate becomes blue color under HRP enzyme-catalyzed, reaction is terminated by the addition of a sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450 nm. The concentration of COPT1 in the samples is then determined by comparing the O.D. of the samples to the standard curve.
Other Notes	Small volumes of COPT1 elisa kit vial(s) may occasionally become entrapped in the seal of the product vial during shipment and storage. If necessary, briefly centrifuge the vial on a tabletop centrifuge to dislodge any liquid in the container`s cap. Certain products may require to ship with dry ice and additional dry ice fee may apply.
Searchable Terms for COPT1 purchase	MBS065175 is a ready-to-use microwell, strip plate ELISA (enzyme-linked immunosorbent assay) Kit for analyzing the presence of the Copper Transporter 1 (COPT1) ELISA Kit target analytes in biological samples. The concentration gradients of the kit standards or positive controls render a theoretical kit detection range in biological research samples containing COPT1. The ELISA analytical biochemical technique of the MBS065175 kit is based on COPT1 antibody-COPT1 antigen interactions (immunosorbency) and an HRP colorimetric detection system to detect COPT1 antigen targets in samples. The ELISA Kit is designed to detect native, not recombinant, COPT1. Appropriate sample types may include undiluted body fluids and/or tissue homogenates, secretions. Quality control assays assessing reproducibility identified the intra-assay CV (%) and inter-assay CV(%).

ADDITIONAL DETAILS

NCBI/Uniprot data below describe general gene information for COPT1. It may not necessarily be applicable to this product.

Protein Family	Copper transporter
Precautions	All of MyBioSource's Products are for scientific laboratory research purposes and are not for diagnostic, therapeutics, prophylactic or in vivo use. Through your purchase, you expressly represent and warrant to MyBioSource that you will properly test and use any Products purchased from MyBioSource in accordance with industry standards. MyBioSource and its authorized distributors reserve the right to refuse to process any order where we reasonably believe that the intended use will fall outside of our acceptable guidelines.
Disclaimer	While every efforts were made to ensure the accuracy of the information provided in this datasheet, MyBioSource will not be liable for any omissions or errors contained herein. MyBioSource reserves the right to make changes to this datasheet at any time without prior notice. It is the responsibility of the customer to report product performance issues to MyBioSource within 30 days of receipt of the product. Please visit our Terms & Conditions page for more information.
Product Note	Select online data sheet information is drawn from bioinformatics databases, occasionally resulting in ambiguous or non-relevant product information. It is the responsibility of the customer to review, verify, and evaluate the information to make sure it matches their requirements before purchasing the kit. Our ELISA Kit assays are dynamic research tools and sometimes they may be updated and improved. If the format of this assay is important to you then please request the current manual or contact our technical support team with a presales inquiry before placing an order. We will confirm the current details of the assay. We cannot guarantee the sample manual posted online is the most current manual, it is intended to serve as an example only. Please refer to the instructions For Use provided with the assay kit for precise details.

For Research Use Only. Not for use in diagnostic or therapeutic procedures.

MyBioSource | Tel: +1.858.633.0165 | Fax: +1.858.633.0166 | mybiosource.com/contact-us
Online datasheet: https://www.mybiosource.com/copt1-human-elisa-kits/copper-transporter-1/65175